

MDPRO6000

Patient Monitor

The 6000 series is a sleek and modern designed unit to provide comprehensive monitoring capabilities for patients in various healthcare settings. The MDPro6000 features a high-resolution color display that presents clear and detailed information, allowing for easy interpretation of the essential vital signs. With its user-friendly interface, portability, and robust functionality, the MDPro6000 patient monitor is an essential tool in modern healthcare facilities, enhancing patient safety and improving overall clinical outcomes.



1200

NIBP
Measurement

240 HR

Trend
Review

200

Alarm
Review

48 H

Frozen
Waveform

Features

- High-precision vital signs monitoring with extensive data storage
- Streamline connectivity through our easy data transfer options
- Advanced patient monitoring algorithm with alerts and notifications
- Semi modular design for flexible configuration based on clinical needs
- User-friendly interface for easy operation
- Accessories for all patient types
- G2 CO2 water traps can be used with generic male luer-lock cannula

Standard Parameters: 3/5 lead ECG, NIBP, Spo2, 2-Temp, IBP, RESP, PR and HR

Standard Features: Touch screen, WiFi, USB, 12-inch screen, VGA output, 8GB internal memory, Dual IBP slots

Optional Configurations & Features: 6/12 lead ECG, G2 CO2, Cardiac Output, Thermal Recorder, Nurse Call (with CMS), Defibrillator Synchronization

12" Touch Screen



Primary Care



ASC




Dental

Proprietary Algorithms & Technologies



G2 CO2 (sidestream)

- Superior water trap design for accurate monitoring
- iCARBTM algorithm with Intelligent CO2 pseudo wave identification technology
- Sampling rate as low as 50ml/min



ECG

- 12-lead ST analysis optional with additional internal module upgrade
- Customizable 3/5-lead placement for more/ less ECG waves
- Automatic lead type detection
- Industry leading iSEAPTTM algorithm with auto-detection of 33 types of arrhythmias.
- SEMIP[®] algorithm with 208 ECG findings over age/gender diversities

NIBP

- Dual dust filter design means no blockage inside and provides accurate NIBP readings
- Unique cleaning mode for routine maintenance
- iCUFSTM algorithm with smart deflation technology



SpO2

- iMAT algorithm with motion resistance and low perfusion resistance performance
- Reference reading of Perfusion Index (PI) from 0 to 10 according to perfusion changes
- Simultaneous measurements of SpO2 and NIBP of the same limb

Configurations

MDPro6000

Standard Configuration with WiFi & Touch Screen

MDPro6000.CO

Standard Configuration with WiFi, Touch Screen and Cardiac Output

MDPro6000-G2

Standard Configuration with internal CO₂, WiFi & Touch Screen with internal OEM MDPro Sidestream CO₂. Uses traditional water traps and generic cannulas

MDPro6000.P

Standard Configuration with WiFi, Touch Screen & Thermal Printer

MDPro6000.CO.P

Standard Configuration with WiFi, Touch Screen, Cardiac Output & Thermal Printer

MDPro6000-G2.P

Standard Configuration with internal CO₂, WiFi, Touch Screen & Thermal Printer

Accessories

STANDARD ACCESSORIES

- ECG cable, 3-lead, snap, AHA, 3.4m — **01.57.471388**
- SpO₂ Finger Sensor, Adult, 2.5m, reusable - direct connect 7 pin — **02.57.225029**
- NIBP Cuff, Adult, 27cm-35cm, reusable — **Cuff.E9**
- NIBP Tube — **01.59.473007**
- Adult skin temperature probe — **01.15.040225**
- Rechargeable Lithium-Ion Battery (10.8V, 2550mAh) — **01.21.064380**

G2 ACCESSORIES

- Water Trap — **02.01.210520**
- ETCO₂ Sampling Cannulas, Adult cannula with 7' CO₂ line. Male Luer-Lok Connector — **4000-7-25**
- ETCO₂ Sampling Lines 10' (Male to Female) — **4410-10-25**

Optional Accessories

SPO₂ SENSORS

- SpO₂ Finger Sensor, Adult, 2.5m, reusable — **SH1.DB9**
- SpO₂ Warp Sensor, Neonate, 1m, reusable — **SH3.DB9**
- SpO₂ Silicone Soft-tip Sensor, Adult, 1m, reusable — **SH4.DB9**
- SpO₂ Silicone Soft-tip Sensor, Pediatric, 1m, reusable — **SH5.DB9**
- SpO₂ Ear Clip Sensor, Adult/Pediatric, 1m, reusable — **SH6.DB9**
- SpO₂ 7-pin Extension Cable, 2m — **01.57.471068**
- SpO₂ 7-pin Extension Cable, 4m — **01.57.471789**

CUFFS

- NIBP Cuff, Infant, 10-15cm, reusable — **Cuff.E5**
- NIBP Cuff, Small Child, 13-17cm, reusable — **Cuff.E6**
- NIBP Cuff, Child, 16-21cm, reusable — **Cuff.E7**
- NIBP Cuff, Small Adult, 20.5-28cm, reusable — **Cuff.E8**
- NIBP Cuff, Adult, 27cm-35cm, reusable — **Cuff.E9**
- NIBP Cuff, Large Adult, 34cm-43cm, reusable — **Cuff.E10**

NIBP TUBING

- NIBP Tube (3m) with connector — **01.59.036118-11**

Specifications

Physical Specification

Device Dimension:
261 mm (W)×246 mm (H)×146 mm (D) Weight:
approx. < 2.8 kg

Display

Color TFT LCD: 12"
Resolution: 800×480
Waveforms Displayed: Up to 13

Lead Mode: 3 Electrodes: I, II, III
5 Electrodes: I, II, III, aVR, aVL, aVF, V
6 Electrodes: I, II, III, aVR, aVL, aVF, and leads
corresponding to Va Vb.
10 Electrodes: I, II, III, aVR, aVL, aVF, V1, V2, V3, V4, V5,

Sweep Speed: 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50
mm/s
CMRR: Diagnosis: > 95 dB
Diagnosis I: > 105 dB (when Notch is turned on)
Monitor: > 105 dB
Surgery: > 105 dB
Enhanced: > 105 dB
Customized: > 105 dB (Low-pass Filter < 40 Hz) >95 dB
(Low-pass Filter > 40 Hz) Sampling Frequency: 1000

Range:
ADU: 15 bpm to 300 bpm
PED/NEO: 15 bpm to 350 bpm Accuracy: ±1% or 1 bpm,
whichever is greater
Resolution: 1 bpm
Sensitivity: ≥ 300 μV/PP

SPO2

Measuring Range: 0% to 100% Resolution: 1%
Data Update Period: 1 s
Accuracy:
Adult /Pediatric 2% (70% to 100% SpO2) Undefined:
(0% to 69% SpO2) Neonate: 3% (70% to 100% SpO2)
Undefined: (0% to 69% SpO2)
Sensor:
Red Light (660±/3) nm I Infrared Light
+/-10) nm Emitted Light Energy: < 15 mW

Measuring Range: 0-10, invalid PI value is 0.
Resolution: 1

Method:
Impedance between RA-LL, RA-LA Measurement

Options are lead I and II. The default is Lead II.
Calculation Type: Manual, Automatic Baseline
Impedance Range: 200 Ω to 2500 Ω (with ECG
cables of 1 kΩ resistance)
Measuring Sensitivity: Within the baseline
impedance range: 0.3 Ω Waveform Bandwidth: 0.2
Hz to 2.5 Hz (-3 dB)
Respiration Excitation Waveform: Sinusoid, 45.6 kHz
(10%), < 350 μA RR Measuring Range:

Adult: 0 rpm to 120 rpm Neo/Ped0
rpm to 150 rpm Resolution 1 rpm

Accuracy:

Adult: 6 rpm to 120 rpm: 2 rpm 0 rpm to 5 rpm: not
specified

Neo/Ped6 rpm to 150 rpm: 2 rpm 0 rpm to 5 rpm: not
specified Gain Selection: 0.25, 0.5, 1, 2, 3, 4, 5 Sweep:
6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
No RR Detected Delay: 10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40
s; default value is 20 s.

TEMP

Technique: Thermal resistance Position: Skin, oral
cavity, rectum Measure Parameter: T1, T2, TD(the
absolute value of T2 minus T1) Channel: 2
Sensor Type: YSI-10K and YSI-2.252K Unit: °C, °F
Measuring Range: 0 °C to 50 °C (32 °F to 122 °F)
Resolution: 0.1 °C (0.1 °F)
Accuracy: 0.3 °C
Refresh Time: Every 1 s to 2 s Temperature
Calibration: At an interval of 5 to 10 minutes
Measuring Mode: Direct Mode Transient Response
Time: ≤ 30 s

NIBP

Technique: Oscillometry
Mode: Manual, Auto, Continuous, Sequence
Measuring Interval in AUTO Mode (unit:
minutes): 1/2/3/4/5/10/15/30 /60/90/120/180/240/360/
480 and User Define
Continuous: 5 min, interval is 5 s Measuring
Parameter: SYS, DIA, MAP, P R
Pressure Unit: kPa, mmHg, cmH2O Measuring
Range:

Adult Mode:

SYS: 25 mmHg to 290 mmHg DIA: 10
mmHg to 250 mmHg MAP: 15 mmHg to
260 mmHg Pediatric Mode:
SYS: 25 mmHg to 240 mmHg DIA: 10
mmHg to 200 mmHg MAP: 15 mmHg to
215 mmHg Neonatal Mode:
SYS: 25 mmHg to 140 mmHg DIA: 10
mmHg to 115 mmHg MAP: 15 mmHg to
125 mmHg

Alarm Type: SYS, DIA, MAP, PR
(NIBP)
Cuff Pressure Measuring Range: 0 mmHg to 300
mmHg
Pressure Resolution: 1 mmHg Maximum Mean Error:
±5 mmHg Maximum Standard Deviation: 8 mmHg
Maximum Measuring Period:
Adult/Pediatric: 120 s
Neonate: 90 s
Typical Measuring Period: 20 s to 35 s
(depend on HR/motion disturbance)

IBP

Complies with IEC 60601-2-34: 2011.
Technique Direct invasive measurement
Channel 2 channels
IBP
Measure
Measuring
Range
Art (0 to +300) mmHg
PA/PAWP (-6 to +120) mmHg
CVP/RAP/LAP/ICP (-10 to +40) mmHg
P1/P2 (-50 to +300) mmHg
Resolution 1 mmHg
Accuracy (not including sensor) ± 2 % or ±1 mmHg,
whichever is greater
ICP:
0 mmHg to 40 mmHg: ± 2 % or ±1
mmHg, whichever is greater;
-10 mmHg to -1 mmHg: undefined
Pressure Unit kPa, mmHg, cmH2O
Pressure sensor
Sensitivity 5 μV/V/mmHg
Impedance Range 300 Ω to 3000 Ω
Filter DC~ 12.5 Hz; DC~ 40 Hz
Zero Range: ± 200 mmHg
Pressure Calibration
Range
IBP (excluding ICP) 80 mmHg to 300 mmHg
ICP 10 mmHg to 40 mmHg
Volume Displacement 7.4 x 104 mm3 / 100 mmHg

CO2

Complies with ISO 80601-2-55: 2011.
Intended Patient Adult, pediatric, neonatal
Measure Parameters EtCO2, FiCO2, AwRR
Unit mmHg, %, kPa
Measuring
Range
EtCO2 0 mmHg to 150 mmHg (0 % to 20%)
FiCO2 0 mmHg to 50 mmHg
AwRR 2 rpm to 150 rpm
Resolution
EtCO2 1 mmHg
FiCO2 1 mmHg
AwRR 1 rpm
Accuracy EtCO2
± 2 mmHg, 0 mmHg to 40
mmHg Typical conditions:
Ambient temperature: (25 ± 3) °C
Barometric pressure: (760 ± 10) mmHg
Balance gas: N2
Sample gas flowrate: 100 ml/min
± 5% of reading, 41 mmHg to 70 mmHg
± 8% of reading, 71 mmHg to 100 mmHg
± 10% of reading, 101 mmHg to 150 mmHg
± 12% of reading or ± 4 mmHg, whichever is greater All
conditions AwRR ± 1 rpm
Drift of Measure Accuracy Meets the requirements of the
measure accuracy
Sample Gas Flowrate 70 ml/min or 100 ml/min (default),
accuracy: ±15 ml/min
Warm-up Time Display reading within 20 s; reach to the
designed accuracy within 2 minutes.
Rise Time < 400 ms (with 2 m gas sampling tube, sample
gas flowrate: 100 ml/min) < 500 ms (with 2 m gas
sampling tube, sample gas flowrate: 70 ml/min)
Response Time < 4 s
(with 2 m gas sampling tube, sample gas flowrate: 100
ml/min/70 ml/min)
Work Mode Standby (default), measure O2
Compensation Range: 0% to 100%
Resolution: 1%
Default: 16%
N2O
Compensation
Range: 0% to 100%
Resolution: 1%
Default: 0%
AG
Compensation
Range: 0% to 20%
Resolution: 0.1%
Default: 0%
Humidity Compensation Method
ATPD (default), BTPS
Barometric Pressure Compensation Automatic
(The change of barometric pressure will not add
additional errors to the measurement values.)
Zero Calibration Support
Calibration Support (It is recommend to be operated by
trained personal.)
Alarm EtCO2, FiCO2, AwRR
No RR
Detected
Delay
10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s; default value is 20 s.
Data Sample
Rate
100 Hz
EtCO2
Change1
AwRR ≤ 80 rpm, meet the accuracy mentioned above;
AwRR > 80 rpm, EtCO2 descends 8%;
AwRR > 90 rpm, EtCO2 descends 10%;
with 2 m gas sampling tube, sample gas flowrate: 100 ml/
min)
AwRR ≤ 60 rpm, meet the accuracy mentioned above;
AwRR > 60 rpm, EtCO2 descends 8%;
AwRR > 90 rpm, EtCO2 descends 10%;
AwRR > 120 rpm, EtCO2 descends 15%;
with 2 m gas sampling tube, sample gas flowrate: 70 ml/
min)